



399 Reservation Road
Marina, CA 93933 USA
Phone: 831-384-4300
Fax: 831-384-4312
sales@foxthermal.com

FT4X CAL-V™ CERTIFICATE CALIBRATION VALIDATION

CAL-V™ Performed on: August 18 2022 2:32:51 PM

Firmware version: FT4X V7.3

Fox Meter Serial Number: ENG100

CAL-V™ Results: Pass

CAL-V™: 0.05

Test Temperature 77.3 F

Tag #/Meter Location: 12345

Test performed by: John Doe

Additional Comments: Well #3

CAL-V™ is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Repeatability of sensor
- * Repeatability of sensor electronics
- * Confirms Calibration Algorithms

At the conclusion of the test, the meter will display a pass/fail message and the CAL-V™ data.

A "pass" result confirms the meter is measuring accurately.

CAL-V™ limits: $\pm 0 - 0.8$ Pass, $\pm 0.8-1.0$ Warning, $> \pm 1.0$ Fail

Configuration:

Pipe Diameter:	9 In	Gas SelectX:	Mix Gas
Customer STP:	32.0 Deg F & 760.00 mmHG	CO2	40%
4-20 mA Range:	0 - 5001 SCFM	Methane	56%
Zero Flow Cutoff:	1.6768366 SCFM	Nitrogen	1%
Previous CAL-V:	0.51	Oxygen	2%
Previous CAL-V:	Pass	Hydrogen	1%
Gross Heating Value[BTU/FT3]:	605		
Density[Kg/M3]:	1.2332		