

CERTIFICATE OF CONFORMITY

1. HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT PER US REQUIREMENTS

2. **Certificate No:** FM17US0061X

3. **Equipment:**

(Type Reference and Name)

Model FT4X Thermal Mass Flow Meter and Temperature

Transmitter

4. Name of Listing Company: Fox Thermal Instruments Inc.

Address of Listing Company: 5.

399 Reservation Road Marina, CA 93933

USA

6. The examination and test results are recorded in confidential report number:

3061039 dated 31st March 2017

7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:

FM Class 3600:2011. FM Class 3615:2006. FM Class 3810:2005. ANSI/ISA 60079-0:2013, ANSI/ISA 60079-1:2013, ANSI/NEMA 250:1991, ANSI/IEC 60529:2004

- If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.
- This certificate relates to the design, examination and testing of the products specified herein. The FM 9. Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.
- 10. Equipment Ratings:

Explosion proof for Class I, Division 1, Groups B, C and D; Dust-ignition proof for Class II, Division 1, Groups E, F and G: Class III. Division 1: Flameproof for Class I. Zone 1. AEx db IIB+H2 T6 or T4 Gb hazardous (classified) locations, indoors and outdoors (Type 4X, IP67) with an ambient temperature rating of -40°C to +70°C (-20°C for Zones).

Certificate issued by:

E. Marquestion

31 March 2017

Date

J. E. Marquedant Manager, Electrical Systems

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

Page 1 of 3 F 347 (Mar 16)

SCHEDULE



Member of the FM Global Group

US Certificate Of Conformity No: FM17US0061X

11. The marking of the equipment shall include:

Main Enclosure:

Class I Division 1, Groups B, C, D; T6 or T4 Ta = -40°C to +70°C; Type 4X, IP67

Class II, Division 1, Groups E, F, G, Class III, Division 1; T6 or T4 Ta = -40°C to +70°C; Type 4X, IP67

Class I, Zone 1, AEx db IIB+H2 T6 or T4 Gb Ta = -20°C to +70°C, Type 4X, IP67

Remote Enclosure:

Class I Division 1, Groups B, C, D; T4 Ta = -40°C to +70°C; Type 4X, IP67

Class II, Division 1, Groups E, F, G, Class III, Division 1; T4 Ta = -40°C to +70°C; Type 4X, IP67

Class I, Zone 1, AEx db IIB+H2 T4 Gb Ta = -20°C to +70°C, Type 4X, IP67

12. Description of Equipment:

General - The Model FT4X Thermal Mass Flow Meter and Temperature Transmitters are used for measuring the mass flow of gases as well as the process gas temperature all in one package. The flowmeters are available for inline or insertion applications. The product features 4-20mA and/or pulse outputs and is also available with an optional display/configuration panel. Digital communications options include USB, HART and RS485 Modbus.

Construction - The Model FT4X Thermal Mass Flow Meter and Temperature Transmitter consists of a dual-compartment electronics housing with thread-on blank or window covers and a probe threaded into the housing base. The remote version of the Model FT4X Thermal Mass Flow Meter and Temperature Transmitter consists of the same dual-compartment electronics housing attached to a smaller remote enclosure with probe. The equipment enclosures are manufactured by Limatherm. The remote enclosure, Limatherm model XD-AD, is FM certified for Class I, Division 1, Groups A, B, C and D and for Class II, Division 1, Groups E, F and G, Class III and for Class I, Zone 1, AEx d IIC Type 4X. The main enclosure, Limatherm models XD-ID100 or XD-ID100win, is FM certified for Class I, Division 1, Groups A, B, C and D and for Class II, Division 1, Groups E, F and G, Class III and for Class I, Zone 1, AEx db IIC Gb, Type 4X. The XD-ID100 enclosure is rated for Ta = -40°C to +100°C (or +85°C for "win"). The enclosures are provided with ¾ inch NPT openings. Probe mounting can be provided integrally, directly to the electronics housing, or can be remotely mounted, using the remote enclosure, up to 100 feet from the electronics housing.

Ratings - The Model FT4X Thermal Mass Flow Meter and Temperature Transmitters operate at 12-28 Vdc (6W). The transmitters are rated for use in an ambient temperature range of -40°C to +70°C for Divisions or -20°C to +70°C for Zones. The transmitter probes are rated for use in a process temperature range of -40°C to +121°C at a maximum pressure of 500 psi.

FT4X-a-b-c-d.Thermal Mass Flow Meter and Temperature Transmitter (Insertion).

- a = Probe 04I, 05I, 06I, 07I, 08I, 09I, 10I, 11I, 12I, 13I, 14I, 15I, 16I, 17I, 18I, 19I, 20I, 21I, 22I, 23I, 24I, 25I, 26I, 27I, 28I, 29I, 30I, 31I, 32I, 33I, 34I, 35I, 36I, 15R, 18R, 24R, 30R or 36R.
- b = Enclosure E1 or E3.
- c = Communication BH, BN, B0 or MB.
- d = Data logger and display L0, L1, D0 or LD0.

FT4X-a-b-c-d.Thermal Mass Flow Meter and Temperature Transmitter (Inline).

a = Flow body 05F, 075F, 10F, 15F, 20F, 25F, 30F, 40F, 60F, 125F, 20FC, 25FC, 30FC, 40FC, 075G, 10G, THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

F 347 (Mar 16) Page 2 of 3

SCHEDULE



Member of the FM Global Group

US Certificate Of Conformity No: FM17US0061X

15G, 20G, 25G, 30G, 40G, 60G, 125G, 025P, 05P, 075P, 10P, 15P, 20P, 25P, 30P, 40P, 125P, 20PC, 25PC, 30PC or 40PC.

- b = Enclosure E1 or E3.
- c = Communication BH, BN, B0 or MB.
- d = Data logger and display L0, L1, D0 or LD0.

13. Specific Conditions of Use:

- 1. The flamepaths of the equipment are not intended to be repaired. Consult the manufacturer if repair of the flamepath joints is necessary.
- 2. Refer to the manufacturer's instructions to reduce the potential of an electrostatic charging hazard on the equipment enclosure.
- 3. The equipment temperature code ratings are dependent on the enclosure configuration (local or remote). Refer to the following table for specific temperature code markings.

Model Code	Temperature Code Marking		
Enclosure (variable "b")	Main Enclosure	Remote Enclosure	
E1	T4	N/A	
E3	T6	T4	

14. Test and Assessment Procedure and Conditions:

This Certificate has been issued in accordance with FM Approvals US Certification Requirements.

15. Schedule Drawings

A copy of the technical documentation has been kept by FM Approvals.

16. Certificate History

Details of the supplements to this certificate are described below:

Date	Description	N. I	
31st March 2017	Original Issue.	Annrough	0

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

F 347 (Mar 16) Page 3 of 3