



399 Reservation Road
Marina, CA 93933 USA
Phone: 831-384-4300
Fax: 831-384-4312
sales@foxthermal.com

CAL-V™ CERTIFICATE CALIBRATION VALIDATION

CAL-V™ Performed on: May 15 2020 7:42:50 AM

Firmware version: FT1 V6.0

Fox Meter Serial Number: -----

CAL-V™ Results: CAL-V PASS

CAL-V™ Value: -0.07

Test Temperature 70.2 F

Tag #/Meter Location: Tag

Test performed by: Patrick

Additional Comments: Some Comments

CAL-V™ is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Sensor free of residue or film
- * Zero Stability of the meter
- * Thermal Conductivity (heat transfer) repeatability of the sensor

At the conclusion of the test, the meter will display a pass/fail message and the CAL-V™ data.

A "pass" result confirms the meter is measuring accurately.